

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
 Data File : VN068924.D
 Acq On : 14 Oct 2021 15:00
 Operator : JC/MD
 Sample : M4192-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HAR-6

Quant Time: Oct 14 22:45:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	377738	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	619288	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	569523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	223743	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	223130	44.802	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.600%
35) Dibromofluoromethane	8.024	113	182969	50.781	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.560%
50) Toluene-d8	10.443	98	749983	52.023	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.040%
62) 4-Bromofluorobenzene	12.734	95	258749	47.571	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.140%
Target Compounds						
					Qvalue	
16) Acetone	4.301	43	21142	11.375	ug/l	97
20) Methylene Chloride	5.098	84	23742	5.550	ug/l	91
25) 2-Butanone	7.348	43	5541	1.972	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
Data File : VN068924.D
Acq On : 14 Oct 2021 15:00
Operator : JC/MD
Sample : M4192-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HAR-6

Quant Time: Oct 14 22:45:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

