

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
 Data File : VN068926.D
 Acq On : 14 Oct 2021 15:50
 Operator : JC/MD
 Sample : M4192-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HAR-10

Quant Time: Oct 14 22:46:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	383679	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	626814	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	590862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	235491	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	229502	45.368	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 90.740%			
35) Dibromofluoromethane	8.024	113	187432	51.395	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.800%			
50) Toluene-d8	10.443	98	771573	52.878	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 105.760%			
62) 4-Bromofluorobenzene	12.733	95	272828	49.558	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 99.120%			
Target Compounds						
16) Acetone	4.296	43	18057	9.565	ug/l	98
20) Methylene Chloride	5.103	84	23410	5.388	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
Data File : VN068926.D
Acq On : 14 Oct 2021 15:50
Operator : JC/MD
Sample : M4192-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HAR-10

Quant Time: Oct 14 22:46:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

