

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
 Data File : VN068928.D
 Acq On : 14 Oct 2021 16:41
 Operator : JC/MD
 Sample : M4192-14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HAR-14

Quant Time: Oct 14 22:47:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	383766	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	628063	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	584787	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	228750	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	228920	45.243	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.480%
35) Dibromofluoromethane	8.024	113	189010	51.725	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.460%
50) Toluene-d8	10.443	98	769981	52.664	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.320%
62) 4-Bromofluorobenzene	12.733	95	263079	47.692	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.380%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	21745	11.516	ug/l	98
20) Methylene Chloride	5.103	84	23970	5.515	ug/l	94
25) 2-Butanone	7.343	43	4969	1.740	ug/l	99
43) Isopropyl Acetate	8.571	43	59480	6.084	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
Data File : VN068928.D
Acq On : 14 Oct 2021 16:41
Operator : JC/MD
Sample : M4192-14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HAR-14

Quant Time: Oct 14 22:47:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

