

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
 Data File : VN068932.D
 Acq On : 14 Oct 2021 18:21
 Operator : JC/MD
 Sample : M4192-22
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HAR-21

Quant Time: Oct 14 22:49:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	362974	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	589146	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	555654	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	223413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	213096	44.528	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.060%
35) Dibromofluoromethane	8.024	113	175455	51.187	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.380%
50) Toluene-d8	10.443	98	717513	52.317	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.640%
62) 4-Bromofluorobenzene	12.733	95	257963	49.853	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.700%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	17706	9.914	ug/l	99
18) Methyl Acetate	4.867	43	11145	1.645	ug/l	95
20) Methylene Chloride	5.103	84	22471	5.467	ug/l	95
43) Isopropyl Acetate	8.576	43	71678	7.817	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
Data File : VN068932.D
Acq On : 14 Oct 2021 18:21
Operator : JC/MD
Sample : M4192-22
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HAR-21

Quant Time: Oct 14 22:49:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

