

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
 Data File : VN068933.D
 Acq On : 14 Oct 2021 18:47
 Operator : JC/MD
 Sample : M4192-01
 Misc : 5.45g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HAR-1

Quant Time: Oct 14 22:50:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	402038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	655028	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	571667	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	254138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	235839	44.492	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	88.980%
35) Dibromofluoromethane	8.018	113	193618	50.805	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.600%
50) Toluene-d8	10.443	98	792992	52.005	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.000%
62) 4-Bromofluorobenzene	12.733	95	274048	47.635	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.280%
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.867	43	64955	8.657	ug/l	97
43) Isopropyl Acetate	8.552	43	97779	9.590	ug/l #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
Data File : VN068933.D
Acq On : 14 Oct 2021 18:47
Operator : JC/MD
Sample : M4192-01
Misc : 5.45g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HAR-1

Quant Time: Oct 14 22:50:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

