

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
 Data File : VN068935.D
 Acq On : 14 Oct 2021 19:37
 Operator : JC/MD
 Sample : M4192-05
 Misc : 5.66g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HAR-6

Quant Time: Oct 14 22:50:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	364955	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	600417	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	547021	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	237541	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	216637	45.022	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.040%
35) Dibromofluoromethane	8.019	113	176635	50.564	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.120%
50) Toluene-d8	10.440	98	736700	52.707	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.420%
62) 4-Bromofluorobenzene	12.733	95	256513	48.643	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.280%
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.865	43	35660	5.236	ug/l #	70
20) Methylene Chloride	5.071	84	4143	1.002	ug/l #	85
43) Isopropyl Acetate	8.563	43	37046	3.964	ug/l #	55
95) Naphthalene	15.507	128	43848	7.486	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101421\
Data File : VN068935.D
Acq On : 14 Oct 2021 19:37
Operator : JC/MD
Sample : M4192-05
Misc : 5.66g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HAR-6

Quant Time: Oct 14 22:50:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

