

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101424\
 Data File : VN084409.D
 Acq On : 14 Oct 2024 13:03
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Oct 15 01:34:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 09 02:22:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	26871	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	124285	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	116895	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	66798	30.877	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.933%
60) 4-Bromofluorobenzene	12.847	95	46329	24.014	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	80.033%
63) Toluene-d8	10.565	98	156341	28.870	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.233%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101424\
Data File : VN084409.D
Acq On : 14 Oct 2024 13:03
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Oct 15 01:34:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Oct 09 02:22:28 2024
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 7.812 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN084409.D
Acq: 14 Oct 2024 13:03




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion:128 Resp: 26871
Ion Ratio Lower Upper
128 100
49 171.6 0.0 430.0
130 134.0 0.0 329.8



#27
1,2-Dichloroethane-d4
Concen: 30.877 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. -0.006 min
Lab File: VN084409.D
Acq: 14 Oct 2024 13:03

Tgt Ion: 65 Resp: 66798
Ion Ratio Lower Upper
65 100
67 52.2 41.5 62.3









#28
1,4-Difluorobenzene
Concen: 30.000 ug/l
RT: 9.100 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN084409.D
Acq: 14 Oct 2024 13:03




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion:114 Resp: 124285
Ion Ratio Lower Upper
114 100
63 21.0 17.4 26.0
88 16.4 13.2 19.8



#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN084409.D
Acq: 14 Oct 2024 13:03

Tgt Ion:117 Resp: 116895
Ion Ratio Lower Upper
117 100
82 56.8 46.2 69.4
119 32.5 25.4 38.2









#60
4-Bromofluorobenzene
Concen: 24.014 ug/l
RT: 12.847 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN084409.D
Acq: 14 Oct 2024 13:03

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Tgt Ion: 95 Resp: 46329
Ion Ratio Lower Upper
95 100
174 76.3 58.2 87.2
176 72.6 56.4 84.6



#63
Toluene-d8
Concen: 28.870 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN084409.D
Acq: 14 Oct 2024 13:03

Tgt Ion: 98 Resp: 156341
Ion Ratio Lower Upper
98 100
100 63.8 52.5 78.7

