

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101518\
 Data File : VN051868.D
 Acq On : 15 Oct 2018 12:27
 Operator : MD\SY
 Sample : J5512-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 001M-WILLETS-PT-BLVD-AUG

Quant Time: Oct 16 01:21:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 12 10:23:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	126912	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	709826	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	582879	30.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
27) 1,2-Dichloroethane-d4		8.03	65	234454	30.36	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.20%	
60) 4-Bromofluorobenzene		12.40	95	210177	25.01	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	83.37%	
63) Toluene-d8		10.09	98	809618	31.99	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	106.63%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
15) Acetone	3.82	58	272480	679.71	ug/l	94
25) Chloroform	7.37	83	208966	16.32	ug/l	97
30) 2-Butanone	6.85	43	10075	5.27	ug/l #	66
41) Bromodichloromethane	9.41	83	13988	1.43	ug/l #	82
58) 4-Methyl-2-Pentanone	9.99	43	20645	5.29	ug/l #	93
62) Toluene	10.16	91	4102164	141.67	ug/l	98
82) p-Isopropyltoluene	13.29	119	107380	3.54	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101518\
Data File : VN051868.D
Acq On : 15 Oct 2018 12:27
Operator : MD\SY
Sample : J5512-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
001M-WILLETS-PT-BLVD-AUG

Quant Time: Oct 16 01:21:43 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 12 10:23:44 2018
Response via : Initial Calibration

