

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101518\
 Data File : VN051870.D
 Acq On : 15 Oct 2018 13:33
 Operator : MD\SY
 Sample : J5510-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 001M-WILLETS-PT-BLVD-OCT

Quant Time: Oct 16 16:07:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 12 10:23:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	130932	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	750204	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	639670	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	241378	30.30	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.00%
60) 4-Bromofluorobenzene	12.40	95	240332	26.05	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	86.83%
63) Toluene-d8	10.09	98	841864	30.31	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.03%

Target Compounds						Qvalue
15) Acetone	3.82	58	751492	1817.06	ug/l	95
30) 2-Butanone	6.84	43	23582	11.67	ug/l #	95
58) 4-Methyl-2-Pentanone	9.99	43	11126	2.60	ug/l #	80
62) Toluene	10.16	91	12135706	381.89	ug/l	98
82) p-Isopropyltoluene	13.29	119	322252	9.67	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101518\
Data File : VN051870.D
Acq On : 15 Oct 2018 13:33
Operator : MD\SY
Sample : J5510-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
001M-WILLETS-PT-BLVD-OCT

Quant Time: Oct 16 16:07:55 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 12 10:23:44 2018
Response via : Initial Calibration

