

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101518\
 Data File : VN051873.D
 Acq On : 15 Oct 2018 14:46
 Operator : MD\SY
 Sample : J5512-01DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 001M-WILLETS-PT-BLVD-SEPD

Quant Time: Oct 16 16:04:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 12 10:23:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	129993	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	736095	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	591568	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	244496	30.91	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery	=	103.03%	
60) 4-Bromofluorobenzene	12.40	95	207932	24.38	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery	=	81.27%	
63) Toluene-d8	10.09	98	829811	32.30	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery	=	107.67%	

Target Compounds

						Qvalue
15) Acetone	3.82	58	50201	122.26	ug/l	91
25) Chloroform	7.37	83	62020	4.73	ug/l	86
58) 4-Methyl-2-Pentanone	9.99	43	5482	1.38	ug/l	96
62) Toluene	10.16	91	705969	24.02	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101518\
Data File : VN051873.D
Acq On : 15 Oct 2018 14:46
Operator : MD\SY
Sample : J5512-01DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
001M-WILLETS-PT-BLVD-SEPDL

Quant Time: Oct 16 16:04:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 12 10:23:44 2018
Response via : Initial Calibration

