

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101518\
 Data File : VN051874.D
 Acq On : 15 Oct 2018 15:11
 Operator : MD\SY
 Sample : J5512-02DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 002M-35TH-AVE-SEPDL

Quant Time: Oct 16 16:05:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 12 10:23:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	126052	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	711219	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	581547	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	241342	31.47	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	104.90%
60) 4-Bromofluorobenzene	12.40	95	202428	24.14	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	80.47%
63) Toluene-d8	10.09	98	812223	32.16	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	107.20%

Target Compounds

					Qvalue
15) Acetone	3.82	58	61156	153.60	ug/l 94
25) Chloroform	7.37	83	44635	3.51	ug/l 93
58) 4-Methyl-2-Pentanone	9.99	43	4548	1.17	ug/l # 89
62) Toluene	10.16	91	824073	28.52	ug/l 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101518\
Data File : VN051874.D
Acq On : 15 Oct 2018 15:11
Operator : MD\SY
Sample : J5512-02DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
002M-35TH-AVE-SEPD

Quant Time: Oct 16 16:05:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 12 10:23:44 2018
Response via : Initial Calibration

