

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101518\
 Data File : VN051878.D
 Acq On : 15 Oct 2018 16:49
 Operator : MD\SY
 Sample : J5510-02DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 002M-35TH-AVE-OCTDL

Quant Time: Oct 16 02:34:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 12 10:23:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	125847	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	695539	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	573546	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	231669	30.25	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.83%
60) 4-Bromofluorobenzene	12.40	95	199499	24.12	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	80.40%
63) Toluene-d8	10.09	98	784832	31.51	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	105.03%

Target Compounds

						Qvalue
15) Acetone	3.82	58	141810	356.74	ug/l	95
62) Toluene	10.16	91	2016224	70.76	ug/l	99
82) p-Isopropyltoluene	13.29	119	41624	1.39	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101518\
Data File : VN051878.D
Acq On : 15 Oct 2018 16:49
Operator : MD\SY
Sample : J5510-02DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
002M-35TH-AVE-OCTDL

Quant Time: Oct 16 02:34:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 12 10:23:44 2018
Response via : Initial Calibration

