

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101519\
 Data File : VN058789.D
 Acq On : 15 Oct 2019 17:44
 Operator : JC/SP
 Sample : K5326-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IDW-SO-101019

Quant Time: Oct 16 05:22:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	375008	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	597374	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	548333	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	227686	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	275540	52.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.96%	
35) Dibromofluoromethane	7.57	113	194588	52.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.48%	
50) Toluene-d8	10.09	98	764675	51.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.64%	
62) 4-Bromofluorobenzene	12.40	95	254597	45.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.16%	
Target Compounds						
16) Acetone	3.80	43	15096	6.946	ug/l	92
20) Methylene Chloride	4.53	84	1026816	226.047	ug/l	98
43) Isopropyl Acetate	8.15	43	94628	9.094	ug/l #	90
73) Isopropylbenzene	12.25	105	24780	1.503	ug/l	99
78) n-propylbenzene	12.60	91	51049	2.657	ug/l	99
85) sec-Butylbenzene	13.17	105	27348	1.721	ug/l	85
95) Naphthalene	15.13	128	2974048	235.311	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101519\
Data File : VN058789.D
Acq On : 15 Oct 2019 17:44
Operator : JC/SP
Sample : K5326-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IDW-SO-101019

Quant Time: Oct 16 05:22:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

