

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN101519\
 Data File : VN058772.D
 Acq On : 15 Oct 2019 10:43
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 16 04:26:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.623	0.646	-3.7	97	0.00
3 P	Chloromethane	0.740	0.745	-0.7	99	0.00
4 C	Vinyl Chloride	0.738	0.762	-3.3#	99	0.00
5 T	Bromomethane	0.449	0.463	-3.1	101	0.00
6 T	Chloroethane	0.455	0.467	-2.6	100	0.00
7 T	Trichlorofluoromethane	0.883	0.927	-5.0	104	0.00
8 T	Diethyl Ether	0.326	0.334	-2.5	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.503	0.504	-0.2	99	0.00
10 T	Methyl Iodide	0.687	0.684	0.4	94	0.00
11 T	Tert butyl alcohol	0.113	0.122	-8.0	107	0.00
12 CM	1,1-Dichloroethene	0.480	0.491	-2.3#	98	0.00
13 T	Acrolein	0.091	0.088	3.3	93	0.00
14 T	Allyl chloride	0.969	0.992	-2.4	99	0.00
15 T	Acrylonitrile	0.286	0.315	-10.1	103	0.00
16 T	Acetone	0.290	0.343	-18.3	112	0.00
17 T	Carbon Disulfide	1.535	1.498	2.4	96	0.00
18 T	Methyl Acetate	0.735	0.793	-7.9	104	0.00
19 T	Methyl tert-butyl Ether	1.715	1.790	-4.4	98	0.00
20 T	Methylene Chloride	0.606	0.571	5.8	97	0.00
21 T	trans-1,2-Dichloroethene	0.537	0.542	-0.9	96	0.00
22 T	Diisopropyl ether	1.884	1.990	-5.6	99	0.00
23 T	Vinyl Acetate	1.520	1.726	-13.6	100	0.00
24 P	1,1-Dichloroethane	1.073	1.096	-2.1	98	0.00
25 T	2-Butanone	0.419	0.473	-12.9	107	0.00
26 T	2,2-Dichloropropane	0.980	0.984	-0.4	99	0.00
27 T	cis-1,2-Dichloroethene	0.626	0.618	1.3	96	0.00
28 T	Bromochloromethane	0.373	0.378	-1.3	98	0.00
29 T	Tetrahydrofuran	0.271	0.297	-9.6	103	0.00
30 C	Chloroform	1.078	1.079	-0.1#	98	0.00
31 T	Cyclohexane	1.027	0.997	2.9	96	0.00
32 T	1,1,1-Trichloroethane	0.919	0.945	-2.8	96	0.00
33 S	1,2-Dichloroethane-d4	0.700	0.700	0.0	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.309	0.313	-1.3	94	0.00
36 T	1,1-Dichloropropene	0.506	0.512	-1.2	96	0.00
37 T	Ethyl Acetate	0.511	0.575	-12.5	102	0.00
38 T	Carbon Tetrachloride	0.493	0.512	-3.9	97	0.00
39 T	Methylcyclohexane	0.587	0.595	-1.4	94	0.00
40 TM	Benzene	1.452	1.488	-2.5	97	0.00
41 T	Methacrylonitrile	0.239	0.263	-10.0	95	0.00
42 TM	1,2-Dichloroethane	0.558	0.564	-1.1	97	0.00
43 T	Isopropyl Acetate	0.871	0.932	-7.0	102	0.00
44 TM	Trichloroethene	0.361	0.363	-0.6	96	0.00
45 C	1,2-Dichloropropane	0.389	0.403	-3.6#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN101519\
 Data File : VN058772.D
 Acq On : 15 Oct 2019 10:43
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 16 04:26:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.246	0.259	-5.3	98	0.00
47 T	Bromodichloromethane	0.502	0.534	-6.4	99	0.00
48 T	Methyl methacrylate	0.395	0.428	-8.4	100	0.00
49 T	1,4-Dioxane	0.006	0.007	-16.7	102	0.00
50 S	Toluene-d8	1.235	1.239	-0.3	93	0.00
51 T	4-Methyl-2-Pentanone	0.489	0.576	-17.8	103	0.00
52 CM	Toluene	0.888	0.910	-2.5#	96	0.00
53 T	t-1,3-Dichloropropene	0.576	0.616	-6.9	99	0.00
54 T	cis-1,3-Dichloropropene	0.611	0.649	-6.2	98	0.00
55 T	1,1,2-Trichloroethane	0.363	0.360	0.8	99	0.00
56 T	Ethyl methacrylate	0.531	0.593	-11.7	99	0.00
57 T	1,3-Dichloropropane	0.619	0.640	-3.4	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.256	0.284	-10.9	99	0.00
59 T	2-Hexanone	0.378	0.446	-18.0	104	0.00
60 T	Dibromochloromethane	0.375	0.400	-6.7	98	0.00
61 T	1,2-Dibromoethane	0.364	0.380	-4.4	98	0.00
62 S	4-Bromofluorobenzene	0.467	0.463	0.9	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.350	0.350	0.0	95	0.00
65 PM	Chlorobenzene	1.015	1.045	-3.0	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.375	0.399	-6.4	98	0.00
67 C	Ethyl Benzene	1.810	1.972	-9.0#	96	0.00
68 T	m/p-Xylenes	0.679	0.727	-7.1	96	0.00
69 T	o-Xylene	0.657	0.689	-4.9	96	0.00
70 T	Styrene	1.077	1.161	-7.8	96	0.00
71 P	Bromoform	0.275	0.311	-13.1	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.621	3.864	-6.7	96	0.00
74 T	N-amyl acetate	1.626	1.805	-11.0	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.193	1.257	-5.4	99	0.00
76 T	1,2,3-Trichloropropane	1.097	1.202	-9.6	106	0.00
77 T	Bromobenzene	0.881	0.896	-1.7	96	0.00
78 T	n-propylbenzene	4.220	4.610	-9.2	96	0.00
79 T	2-Chlorotoluene	2.619	2.675	-2.1	96	0.00
80 T	1,3,5-Trimethylbenzene	3.091	3.319	-7.4	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.409	0.455	-11.2	101	0.00
82 T	4-Chlorotoluene	2.687	2.800	-4.2	96	0.00
83 T	tert-Butylbenzene	2.648	2.825	-6.7	98	0.00
84 T	1,2,4-Trimethylbenzene	3.069	3.346	-9.0	96	0.00
85 T	sec-Butylbenzene	3.490	3.837	-9.9	97	0.00
86 T	p-Isopropyltoluene	3.172	3.453	-8.9	95	0.00
87 T	1,3-Dichlorobenzene	1.624	1.656	-2.0	95	0.00
88 T	1,4-Dichlorobenzene	1.652	1.652	0.0	96	0.00
89 T	n-Butylbenzene	2.933	3.154	-7.5	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN101519\
Data File : VN058772.D
Acq On : 15 Oct 2019 10:43
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 16 04:26:35 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.568	0.621	-9.3	100	0.00
91 T	1,2-Dichlorobenzene	1.583	1.621	-2.4	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.279	0.275	1.4	100	0.00
93 T	1,2,4-Trichlorobenzene	0.966	0.975	-0.9	92	0.00
94 T	Hexachlorobutadiene	0.509	0.513	-0.8	95	0.00
95 T	Naphthalene	2.775	2.899	-4.5	93	0.00
96 T	1,2,3-Trichlorobenzene	0.943	0.949	-0.6	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6