

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101520\
 Data File : VN064147.D
 Acq On : 15 Oct 2020 16:43
 Operator : JC/MD
 Sample : L4408-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP-1

Manual Integrations
 APPROVED

apatel
 10/19/2020 11:26:14 AM

Quant Time: Oct 16 03:21:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 15 01:16:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	298432	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	570941	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	589899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	216654	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	235565	55.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.84%	
35) Dibromofluoromethane	7.55	113	194262	49.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.32%	
50) Toluene-d8	10.06	98	733064	49.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.32%	
62) 4-Bromofluorobenzene	12.38	95	257486	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
Target Compounds						
16) Acetone	3.79	43	7910m	6.274	ug/l	
17) Carbon Disulfide	4.01	76	4825	0.568	ug/l #	92
64) Tetrachloroethene	10.60	164	39272	8.830	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101520\
Data File : VN064147.D
Acq On : 15 Oct 2020 16:43
Operator : JC/MD
Sample : L4408-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP-1

Manual Integrations
APPROVED

apatel
10/19/2020 11:26:14 AM

Quant Time: Oct 16 03:21:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 15 01:16:44 2020
Response via : Initial Calibration

