

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101520\
 Data File : VN064150.D
 Acq On : 15 Oct 2020 17:55
 Operator : JC/MD
 Sample : L4221-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-02-100120

Quant Time: Oct 16 15:19:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 15 01:16:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	280585	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	536439	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	560763	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	205439	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	224582	56.20	ug/l	0.00
Spiked Amount			Recovery	=	112.40%	
35) Dibromofluoromethane	7.55	113	187737	50.56	ug/l	0.00
Spiked Amount			Recovery	=	101.12%	
50) Toluene-d8	10.06	98	688527	49.15	ug/l	0.00
Spiked Amount			Recovery	=	98.30%	
62) 4-Bromofluorobenzene	12.38	95	239381	49.71	ug/l	0.00
Spiked Amount			Recovery	=	99.42%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101520\
Data File : VN064150.D
Acq On : 15 Oct 2020 17:55
Operator : JC/MD
Sample : L4221-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-02-100120

Quant Time: Oct 16 15:19:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 15 01:16:44 2020
Response via : Initial Calibration

