

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101520\
 Data File : VN064154.D
 Acq On : 15 Oct 2020 19:31
 Operator : JC/MD
 Sample : L4417-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Manual Integrations
 APPROVED

apatel
 10/19/2020 11:26:20 AM

Quant Time: Oct 16 04:04:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 15 01:16:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	271802	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	527210	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	536084	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	191648	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	216313	55.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.76%	
35) Dibromofluoromethane	7.55	113	180615	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
50) Toluene-d8	10.06	98	683386	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
62) 4-Bromofluorobenzene	12.38	95	225911	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
Target Compounds						
64) Tetrachloroethene	10.61	164	1838m	0.455	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101520\
Data File : VN064154.D
Acq On : 15 Oct 2020 19:31
Operator : JC/MD
Sample : L4417-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Quant Time: Oct 16 04:04:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 15 01:16:44 2020
Response via : Initial Calibration

Manual Integrations
APPROVED

apatel
10/19/2020 11:26:20 AM

