

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101520\
 Data File : VN064158.D
 Acq On : 15 Oct 2020 21:07
 Operator : JC/MD
 Sample : L4417-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-GENIE

Manual Integrations
 APPROVED

apatel
 10/19/2020 11:26:24 AM

Quant Time: Oct 16 04:19:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 15 01:16:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	280588	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	527650	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	560074	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	209415	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	217729	54.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.96%	
35) Dibromofluoromethane	7.55	113	185131	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
50) Toluene-d8	10.06	98	681368	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
62) 4-Bromofluorobenzene	12.38	95	241123	50.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.82%	
Target Compounds						
16) Acetone	3.79	43	3594m	3.032	ug/l	Qvalue
64) Tetrachloroethene	10.60	164	1079	0.256	ug/l #	67
73) Isopropylbenzene	12.22	105	3811	0.251	ug/l	89
78) n-propylbenzene	12.57	91	7647	0.452	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101520\
Data File : VN064158.D
Acq On : 15 Oct 2020 21:07
Operator : JC/MD
Sample : L4417-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-GENIE

Manual Integrations
APPROVED

apatel
10/19/2020 11:26:24 AM

Quant Time: Oct 16 04:19:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 15 01:16:44 2020
Response via : Initial Calibration

