

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101520\
 Data File : VN064160.D
 Acq On : 15 Oct 2020 21:55
 Operator : JC/MD
 Sample : L4392-17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HACK-MH

Manual Integrations
 APPROVED

apatel
 10/19/2020 11:26:25 AM

Quant Time: Oct 16 04:27:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 15 01:16:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	283047	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	544839	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	585435	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	227545	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	227152	56.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.70%	
35) Dibromofluoromethane	7.55	113	190593	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
50) Toluene-d8	10.06	98	716890	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
62) 4-Bromofluorobenzene	12.38	95	255981	52.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.68%	
Target Compounds						
16) Acetone	3.79	43	4181m	3.497	ug/l	
73) Isopropylbenzene	12.23	105	5628	0.341	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101520\
Data File : VN064160.D
Acq On : 15 Oct 2020 21:55
Operator : JC/MD
Sample : L4392-17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HACK-MH

Manual Integrations
APPROVED

apatel
10/19/2020 11:26:25 AM

Quant Time: Oct 16 04:27:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 15 01:16:44 2020
Response via : Initial Calibration

