

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101521\
 Data File : VN068951.D
 Acq On : 15 Oct 2021 15:56
 Operator : JC/MD
 Sample : M4175-01DL 10X
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30784DL

Quant Time: Oct 18 02:15:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	364230	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	629162	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	581413	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	216802	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	232249	48.363	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.720%		
35) Dibromofluoromethane	8.024	113	182255	49.789	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.580%		
50) Toluene-d8	10.443	98	776666	53.028	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	106.060%		
62) 4-Bromofluorobenzene	12.734	95	266567	48.240	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	96.480%		
Target Compounds						Qvalue
16) Acetone	4.296	43	13287	7.414	ug/l	99
39) Methylcyclohexane	9.459	83	7611	1.100	ug/l	93
52) Toluene	10.505	92	607754	57.726	ug/l	99
84) 1,2,4-Trimethylbenzene	13.369	105	34174	2.360	ug/l	100
95) Naphthalene	15.504	128	6390	4.398	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101521\
Data File : VN068951.D
Acq On : 15 Oct 2021 15:56
Operator : JC/MD
Sample : M4175-01DL 10X
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30784DL

Quant Time: Oct 18 02:15:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

