

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101521\
 Data File : VN068954.D
 Acq On : 15 Oct 2021 17:12
 Operator : JC/MD
 Sample : M4202-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CS-1-DEPTH-SOILPILE-BOTTOM

Quant Time: Oct 18 02:16:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	348263	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	598327	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	570914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	207514	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	222661	48.492	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.980%
35) Dibromofluoromethane	8.024	113	178805	51.364	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.720%
50) Toluene-d8	10.443	98	753792	54.119	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.240%
62) 4-Bromofluorobenzene	12.733	95	261630	49.786	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.580%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	17391	10.149	ug/l	96
20) Methylene Chloride	5.103	84	30700	7.784	ug/l	96
43) Isopropyl Acetate	8.563	43	36245	3.892	ug/l #	89
95) Naphthalene	15.509	128	4971	4.289	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101521\
Data File : VN068954.D
Acq On : 15 Oct 2021 17:12
Operator : JC/MD
Sample : M4202-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CS-1-DEPTH-SOILPILE-BOTTOM

Quant Time: Oct 18 02:16:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

