

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101521\
 Data File : VN068961.D
 Acq On : 15 Oct 2021 20:08
 Operator : JC/MD
 Sample : M4254-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Oct 18 02:17:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	338669	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	586466	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	533949	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	197393	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	215967	48.366	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.740%
35) Dibromofluoromethane	8.024	113	173640	50.889	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.780%
50) Toluene-d8	10.443	98	727981	53.323	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.640%
62) 4-Bromofluorobenzene	12.733	95	250259	48.586	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101521\
Data File : VN068961.D
Acq On : 15 Oct 2021 20:08
Operator : JC/MD
Sample : M4254-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Oct 18 02:17:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

