

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101521\
 Data File : VN068963.D
 Acq On : 15 Oct 2021 20:59
 Operator : JC/MD
 Sample : M4254-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 18 02:18:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	338456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	585892	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	536012	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	197249	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	217529	48.747	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.500%
35) Dibromofluoromethane	8.024	113	175721	51.550	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.100%
50) Toluene-d8	10.443	98	723177	53.023	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.040%
62) 4-Bromofluorobenzene	12.734	95	249509	48.487	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101521\
Data File : VN068963.D
Acq On : 15 Oct 2021 20:59
Operator : JC/MD
Sample : M4254-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 18 02:18:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

