

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101618\
 Data File : VN051881.D
 Acq On : 16 Oct 2018 9:08
 Operator : MD\SY
 Sample : VN1016WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1016WBL01

Quant Time: Oct 17 03:02:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 09:51:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	927131	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1329017	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1140019	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	411255	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	502132	56.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.32%	
35) Dibromofluoromethane	7.59	113	499308	60.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	120.84%	
50) Toluene-d8	10.09	98	1814271	58.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	117.84%	
62) 4-Bromofluorobenzene	12.40	95	512277	48.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.38%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101618\
Data File : VN051881.D
Acq On : 16 Oct 2018 9:08
Operator : MD\SY
Sample : VN1016WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN1016WBL01

Quant Time: Oct 17 03:02:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101118W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 09:51:20 2018
Response via : Initial Calibration

