

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101619\
 Data File : VN058796.D
 Acq On : 16 Oct 2019 13:14
 Operator : JC/SP
 Sample : K5358-04RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20190897-AMENDED-COMP-KM-CA

Quant Time: Oct 17 02:21:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	379145	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	595196	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	532035	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	207267	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	277361	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
35) Dibromofluoromethane	7.57	113	99128	26.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	53.92%	
50) Toluene-d8	10.09	98	736377	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
62) 4-Bromofluorobenzene	12.40	95	239881	43.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.22%	
Target Compounds						
11) Tert butyl alcohol	4.76	59	29323	34.111	ug/l #	87
16) Acetone	3.80	43	31447	14.311	ug/l	97
20) Methylene Chloride	4.53	84	11673	2.542	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101619\
Data File : VN058796.D
Acq On : 16 Oct 2019 13:14
Operator : JC/SP
Sample : K5358-04RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20190897-AMENDED-COMP-KM-CA

Quant Time: Oct 17 02:21:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

