

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101619\
 Data File : VN058798.D
 Acq On : 16 Oct 2019 13:58
 Operator : JC/SP
 Sample : K5326-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IDW-AQ-101019

Quant Time: Oct 17 08:16:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	381498	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	594497	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	518819	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	209478	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	266936	49.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.96%	
35) Dibromofluoromethane	7.57	113	192272	52.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.72%	
50) Toluene-d8	10.08	98	737654	50.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.46%	
62) 4-Bromofluorobenzene	12.40	95	239293	43.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.10%	
Target Compounds						
16) Acetone	3.80	43	60417	27.324	ug/l	99
18) Methyl Acetate	4.31	43	6011	1.072	ug/l	99
25) 2-Butanone	6.81	43	113485	35.509	ug/l	99
39) Methylcyclohexane	9.08	83	1856	0.266	ug/l	93
52) Toluene	10.15	92	3817	0.362	ug/l	96
95) Naphthalene	15.13	128	171239	14.726	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101619\
Data File : VN058798.D
Acq On : 16 Oct 2019 13:58
Operator : JC/SP
Sample : K5326-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IDW-AQ-101019

Quant Time: Oct 17 08:16:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

