

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101619\
 Data File : VN058805.D
 Acq On : 16 Oct 2019 16:33
 Operator : JC/SP
 Sample : K5399-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FA19-JMW6105

Quant Time: Oct 17 08:49:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	387080	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	620935	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	548723	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	214620	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	298137	55.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.04%	
35) Dibromofluoromethane	7.57	113	209403	54.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.20%	
50) Toluene-d8	10.09	98	796154	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
62) 4-Bromofluorobenzene	12.41	95	258769	44.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.14%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.81	96	1835	0.378	ug/l #	37
44) Trichloroethene	8.83	130	1425	0.318	ug/l	95
64) Tetrachloroethene	10.63	164	2514	0.655	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101619\
Data File : VN058805.D
Acq On : 16 Oct 2019 16:33
Operator : JC/SP
Sample : K5399-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FA19-JMW6105

Quant Time: Oct 17 08:49:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

