

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101619\
 Data File : VN058808.D
 Acq On : 16 Oct 2019 17:39
 Operator : JC/SP
 Sample : K5396-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FA19-JBW7345A-9001

Quant Time: Oct 17 04:39:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	386989	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	613180	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	537118	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	198065	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	290039	53.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.06%	
35) Dibromofluoromethane	7.58	113	205445	54.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.48%	
50) Toluene-d8	10.09	98	777026	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
62) 4-Bromofluorobenzene	12.41	95	247279	43.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.26%	
Target Compounds						
16) Acetone	3.80	43	7699	3.433	ug/l	94
27) cis-1,2-Dichloroethene	6.81	96	49124	10.133	ug/l	85
32) 1,1,1-Trichloroethane	7.55	97	18562	2.608	ug/l	99
37) Ethyl Acetate	6.91	43	14000	2.233	ug/l #	76
44) Trichloroethene	8.82	130	151886	34.350	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101619\
Data File : VN058808.D
Acq On : 16 Oct 2019 17:39
Operator : JC/SP
Sample : K5396-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FA19-JBW7345A-9001

Quant Time: Oct 17 04:39:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

