


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN101619\
Data File : VN058791.D
Acq On : 16 Oct 2019 8:54
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Title : SW846 8260
Last Update : Sat Oct 12 00:00:30 2019



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.4	18997	PASS
75	95	30	60	53.1	47179	PASS
95	95	100	100	100.0	88768	PASS
96	95	5	9	6.2	5483	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.1	66672	PASS
175	174	5	9	7.3	4863	PASS
176	174	95	101	97.7	65144	PASS
177	176	5	9	6.5	4264	PASS