

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101620\
 Data File : VN064180.D
 Acq On : 16 Oct 2020 21:29
 Operator : JC/MD
 Sample : L4406-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-101420-A

Manual Integrations
 APPROVED

MMDadoda
 10/19/2020 4:23:03 PM

Quant Time: Oct 17 02:53:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 15 01:16:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	314832	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	668526	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	690033	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	242204	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	281956	62.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.76%	
35) Dibromofluoromethane	7.55	113	223459	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
50) Toluene-d8	10.06	98	876756	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
62) 4-Bromofluorobenzene	12.38	95	290275	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
Target Compounds						
16) Acetone	3.78	43	20759	15.608	ug/l	96
18) Methyl Acetate	4.30	43	6363m	1.907	ug/l	
20) Methylene Chloride	4.51	84	10707m	2.602	ug/l	
43) Isopropyl Acetate	8.13	43	101605	10.707	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101620\
Data File : VN064180.D
Acq On : 16 Oct 2020 21:29
Operator : JC/MD
Sample : L4406-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-101420-A

Manual Integrations
APPROVED

MMDadoda
10/19/2020 4:23:03 PM

Quant Time: Oct 17 02:53:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 15 01:16:44 2020
Response via : Initial Calibration

