

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101620\
 Data File : VN064187.D
 Acq On : 17 Oct 2020 00:15
 Operator : JC/MD
 Sample : L4413-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30688

Manual Integrations
 APPROVED

MMDadoda
 10/19/2020 4:23:11 PM

Quant Time: Oct 17 03:18:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 15 01:16:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.42	168	2517	50.00	ug/l	-0.21
34) 1,4-Difluorobenzene	8.31	114	461	50.00	ug/l	-0.24
63) Chlorobenzene-d5	11.06	117	118	50.00	ug/l	-0.32
72) 1,4-Dichlorobenzene-d4	13.34	152	251772	50.00	ug/l	0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.54	65	3184	88.82	ug/l	-0.45
Spiked Amount 50.000			Recovery	=	177.64%	
35) Dibromofluoromethane	7.34	113	1835	575.09	ug/l	-0.21
Spiked Amount 50.000			Recovery	=	1150.18%	
50) Toluene-d8	10.25	98	814601m	67661.15	ug/l	0.19
Spiked Amount 50.000			Recovery	=	135322.30%	
62) 4-Bromofluorobenzene	12.05	95	134	32.38	ug/l	-0.33
Spiked Amount 50.000			Recovery	=	64.76%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101620\
Data File : VN064187.D
Acq On : 17 Oct 2020 00:15
Operator : JC/MD
Sample : L4413-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30688

Manual Integrations
APPROVED

MMDadoda
10/19/2020 4:23:11 PM

Quant Time: Oct 17 03:18:41 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101420W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 15 01:16:44 2020
Response via : Initial Calibration

