

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101719\
 Data File : VN058832.D
 Acq On : 17 Oct 2019 16:59
 Operator : JC/SP
 Sample : K5393-13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T-9-U3-(28)

Quant Time: Oct 18 04:21:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	376051	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	599508	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	554338	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	204558	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	280586	53.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.60%	
35) Dibromofluoromethane	7.57	113	197693	53.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.78%	
50) Toluene-d8	10.08	98	773745	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
62) 4-Bromofluorobenzene	12.40	95	243653	43.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.94%	
Target Compounds						
16) Acetone	3.80	43	13727	6.298	ug/l #	89
18) Methyl Acetate	4.31	43	6857	1.240	ug/l	91
20) Methylene Chloride	4.53	84	12893	2.830	ug/l	91
43) Isopropyl Acetate	8.15	43	79203	7.584	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101719\
Data File : VN058832.D
Acq On : 17 Oct 2019 16:59
Operator : JC/SP
Sample : K5393-13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T-9-U3-(28)

Quant Time: Oct 18 04:21:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

