

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101719\
 Data File : VN058833.D
 Acq On : 17 Oct 2019 17:21
 Operator : JC/SP
 Sample : K5393-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T-13-V1-(0)

Quant Time: Oct 18 04:24:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	365980	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	584999	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	545073	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	209520	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	269819	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
35) Dibromofluoromethane	7.57	113	196330	54.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.66%	
50) Toluene-d8	10.08	98	755380	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
62) 4-Bromofluorobenzene	12.41	95	242932	44.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.84%	
Target Compounds						
16) Acetone	3.80	43	12522	5.903	ug/l	98
18) Methyl Acetate	4.31	43	6750	1.255	ug/l #	79
20) Methylene Chloride	4.53	84	13623	3.073	ug/l	95
43) Isopropyl Acetate	8.15	43	101373	9.948	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101719\
Data File : VN058833.D
Acq On : 17 Oct 2019 17:21
Operator : JC/SP
Sample : K5393-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T-13-V1-(0)

Quant Time: Oct 18 04:24:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

