

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101719\
 Data File : VN058835.D
 Acq On : 17 Oct 2019 18:05
 Operator : JC/SP
 Sample : K5393-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T-16-G1-(0)

Quant Time: Oct 18 04:32:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 12 00:00:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	361514	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	575561	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	531701	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	200771	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	271947	53.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.46%	
35) Dibromofluoromethane	7.57	113	195260	54.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.84%	
50) Toluene-d8	10.08	98	746421	52.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.00%	
62) 4-Bromofluorobenzene	12.41	95	236528	43.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.90%	
Target Compounds						
16) Acetone	3.80	43	12279	5.860	ug/l	91
18) Methyl Acetate	4.31	43	10121	1.904	ug/l #	78
20) Methylene Chloride	4.52	84	15510	3.542	ug/l	97
43) Isopropyl Acetate	8.15	43	90604	9.037	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101719\
Data File : VN058835.D
Acq On : 17 Oct 2019 18:05
Operator : JC/SP
Sample : K5393-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T-16-G1-(0)

Quant Time: Oct 18 04:32:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 12 00:00:30 2019
Response via : Initial Calibration

