


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN101719\
 Data File : VN058814.D
 Acq On : 17 Oct 2019 8:24
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101319W.M
 Title : SW846 8260
 Last Update : Fri Oct 18 09:12:23 2019



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.0	20042	PASS
75	95	30	60	52.3	47635	PASS
95	95	100	100	100.0	91130	PASS
96	95	5	9	6.0	5446	PASS
173	174	0.00	2	0.7	481	PASS
174	95	50	100	75.2	68554	PASS
175	174	5	9	7.2	4917	PASS
176	174	95	101	96.1	65898	PASS
177	176	5	9	6.6	4326	PASS