

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101722\
 Data File : VN074920.D
 Acq On : 17 Oct 2022 16:56
 Operator : JC\MD
 Sample : VN1017WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1017WBL01

Quant Time: Oct 18 03:02:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101722W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 17 16:03:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	315047	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	569684	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	586558	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	237904	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	403567	48.562	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.120%
35) Dibromofluoromethane	8.177	113	318016	49.650	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.300%
50) Toluene-d8	10.571	98	1234243	46.469	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.940%
62) 4-Bromofluorobenzene	12.853	95	289813	37.557	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	75.120%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101722\
Data File : VN074920.D
Acq On : 17 Oct 2022 16:56
Operator : JC\MD
Sample : VN1017WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1017WBL01

Quant Time: Oct 18 03:02:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101722W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 17 16:03:54 2022
Response via : Initial Calibration

