

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101722\
 Data File : VN074923.D
 Acq On : 17 Oct 2022 18:08
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Oct 18 03:05:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101722W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 17 16:03:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	265028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	506857	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	522297	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	215443	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	389581	55.726	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.460%
35) Dibromofluoromethane	8.177	113	309604	54.329	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.660%
50) Toluene-d8	10.571	98	1146466	48.515	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.020%
62) 4-Bromofluorobenzene	12.847	95	272437	39.682	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.360%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101722\
Data File : VN074923.D
Acq On : 17 Oct 2022 18:08
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Oct 18 03:05:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101722W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 17 16:03:54 2022
Response via : Initial Calibration

