

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101722\
 Data File : VN074926.D
 Acq On : 17 Oct 2022 19:20
 Operator : JC\MD
 Sample : N5149-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-GW-918-920

Quant Time: Oct 18 03:09:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101722W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 17 16:03:54 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/18/2022
 Supervised By : Mahesh Dadoda 10/18/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	293162	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	531158	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	554090	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	220690	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	386963	50.040	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.080%			
35) Dibromofluoromethane	8.177	113	296485	49.646	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.300%			
50) Toluene-d8	10.571	98	1184903	47.847	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.700%			
62) 4-Bromofluorobenzene	12.853	95	291430	40.506	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 81.020%#			
Target Compounds					Qvalue	
16) Acetone	4.477	43	12667	3.484	ug/l	98
18) Methyl Acetate	5.059	43	11745m	1.020	ug/l	
44) Trichloroethene	9.353	130	1581	0.286	ug/l	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101722\
Data File : VN074926.D
Acq On : 17 Oct 2022 19:20
Operator : JC\MD
Sample : N5149-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW5-GW-918-920

Quant Time: Oct 18 03:09:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101722W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 17 16:03:54 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/18/2022
Supervised By :Mahesh Dadoda 10/18/2022

