

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101723\
 Data File : VN079588.D
 Acq On : 17 Oct 2023 17:57
 Operator : JC\MD
 Sample : 04913-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Oct 18 02:33:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 18 02:19:15 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/18/2023
 Supervised By : Mahesh Dadoda 10/20/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	18413m	30.000	ug/l	-0.01
28) 1,4-Difluorobenzene	9.106	114	126567	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	144124	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	82180	31.391	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	104.633%
60) 4-Bromofluorobenzene	12.853	95	52572	19.947	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	66.500%
63) Toluene-d8	10.571	98	178550	28.911	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.367%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101723\
Data File : VN079588.D
Acq On : 17 Oct 2023 17:57
Operator : JC\MD
Sample : 04913-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Oct 18 02:33:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Oct 18 02:19:15 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/18/2023
Supervised By :Mahesh Dadoda 10/20/2023

