


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101723\
Data File : VN079570.D
Acq On : 17 Oct 2023 08:53
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101723W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Wed Oct 18 02:19:15 2023



AutoFind: Averaged scan 1845 to 1861; Bkg corrected with scan 1844

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	23.0	3229	PASS
75	95	30	60	55.7	7834	PASS
95	95	100	100	100.0	14065	PASS
96	95	5	9	6.5	908	PASS
173	174	0.00	2	1.2	131	PASS
174	95	50	100	76.9	10816	PASS
175	174	5	9	7.8	848	PASS
176	174	95	101	96.0	10386	PASS
177	176	5	9	6.2	639	PASS