

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101818\
 Data File : VN051917.D
 Acq On : 18 Oct 2018 15:34
 Operator : MD\SY
 Sample : J5580-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 2-AVE-23

Quant Time: Oct 18 17:25:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 12 10:23:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	158381	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	2124982	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	1801411	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	297306	30.85	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.83%
60) 4-Bromofluorobenzene	12.40	95	277989	10.70	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	35.67%#
63) Toluene-d8	10.09	98	978171	12.51	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	41.70%#

Target Compounds

					Qvalue
22) cis-1,2-Dichloroethene	6.83	96	36327	3.71 ug/l	# 72
25) Chloroform	7.37	83	6711	0.42 ug/l	# 76
37) Trichloroethene	8.84	130	13300	0.53 ug/l	87
61) Tetrachloroethene	10.63	164	52865	2.29 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101818\
Data File : VN051917.D
Acq On : 18 Oct 2018 15:34
Operator : MD\SY
Sample : J5580-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
2-AVE-23

Quant Time: Oct 18 17:25:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 12 10:23:44 2018
Response via : Initial Calibration

