

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101818\
 Data File : VN051920.D
 Acq On : 18 Oct 2018 16:54
 Operator : MD\SY
 Sample : J5580-01RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 2-AVE-23RE

Quant Time: Oct 18 17:19:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 12 10:23:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	130340	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	1922997	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	1650802	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	241652	30.47	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery	=	101.57%	
60) 4-Bromofluorobenzene	12.40	95	216601	9.10	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery	=	30.33%#	
63) Toluene-d8	10.09	98	780247	10.88	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery	=	36.27%#	

Target Compounds

						Qvalue
22) cis-1,2-Dichloroethene	6.84	96	36034	4.48	ug/l	# 72
25) Chloroform	7.37	83	7087	0.54	ug/l	# 82
37) Trichloroethene	8.84	130	14093	0.62	ug/l	86
61) Tetrachloroethene	10.63	164	54107	2.56	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101818\
Data File : VN051920.D
Acq On : 18 Oct 2018 16:54
Operator : MD\SY
Sample : J5580-01RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
2-AVE-23RE

Quant Time: Oct 18 17:19:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N101218W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 12 10:23:44 2018
Response via : Initial Calibration

