

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101818\
 Data File : VN051921.D
 Acq On : 18 Oct 2018 17:21
 Operator : MD\SY
 Sample : PB113908ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113908ZHE#02

Manual Integrations
 APPROVED

MMDadoda
 10/19/2018 4:37:58 PM

Quant Time: Oct 19 07:17:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101718W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 19 05:11:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	992955	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1452967	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1276706	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	446086	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	533658	57.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.94%	
35) Dibromofluoromethane	7.59	113	540011	58.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.94%	
50) Toluene-d8	10.09	98	1971309	58.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.28%	
62) 4-Bromofluorobenzene	12.40	95	558541	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
Target Compounds						
20) Methylene Chloride	4.56	84	15562m	1.88	ug/l	Qvalue
43) Isopropyl Acetate	8.17	43	271091	25.81	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101818\
Data File : VN051921.D
Acq On : 18 Oct 2018 17:21
Operator : MD\SY
Sample : PB113908ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113908ZHE#02

Manual Integrations
APPROVED

MMDadoda
10/19/2018 4:37:58 PM

Quant Time: Oct 19 07:17:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101718W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 19 05:11:28 2018
Response via : Initial Calibration

