

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101818\
 Data File : VN051924.D
 Acq On : 18 Oct 2018 18:41
 Operator : MD\SY
 Sample : PB113908ZHE#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113908ZHE#05

Manual Integrations
 APPROVED

MMDadoda
 10/19/2018 4:37:58 PM

Quant Time: Oct 19 07:23:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101718W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 19 05:11:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	988452	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1443409	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1252925	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	457240	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	534663	58.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.70%	
35) Dibromofluoromethane	7.59	113	527643	57.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.02%	
50) Toluene-d8	10.09	98	1976797	59.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.40%	
62) 4-Bromofluorobenzene	12.40	95	563141	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
Target Compounds						
20) Methylene Chloride	4.56	84	14251m	1.73	ug/l	Qvalue
43) Isopropyl Acetate	8.17	43	288815	27.79	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101818\
Data File : VN051924.D
Acq On : 18 Oct 2018 18:41
Operator : MD\SY
Sample : PB113908ZHE#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113908ZHE#05

Manual Integrations
APPROVED

MMDadoda
10/19/2018 4:37:58 PM

Quant Time: Oct 19 07:23:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101718W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 19 05:11:28 2018
Response via : Initial Calibration

