

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN101818\
 Data File : VN051926.D
 Acq On : 18 Oct 2018 19:35
 Operator : MD\SY
 Sample : PB113908ZHE#07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113908ZHE#07

Quant Time: Oct 19 07:29:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101718W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 19 05:11:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1005961	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1475348	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1303542	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	456616	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	542848	58.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.42%	
35) Dibromofluoromethane	7.59	113	546572	58.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.58%	
50) Toluene-d8	10.09	98	2018266	59.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.26%	
62) 4-Bromofluorobenzene	12.40	95	577903	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.55	84	14809	1.76	ug/l	91
43) Isopropyl Acetate	8.17	43	284233	26.70	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN101818\
Data File : VN051926.D
Acq On : 18 Oct 2018 19:35
Operator : MD\SY
Sample : PB113908ZHE#07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113908ZHE#07

Quant Time: Oct 19 07:29:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101718W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 19 05:11:28 2018
Response via : Initial Calibration

