

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101819\
 Data File : VN058851.D
 Acq On : 18 Oct 2019 12:39
 Operator : JC/SP
 Sample : PB124110TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB124110TB

Quant Time: Oct 21 07:40:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	396609	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	641196	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	602869	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	229985	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	294799	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
35) Dibromofluoromethane	7.57	113	208694	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
50) Toluene-d8	10.08	98	836280	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
62) 4-Bromofluorobenzene	12.40	95	272858	45.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.80%	
Target Compounds						
16) Acetone	3.80	43	13249	4.609	ug/l	96
18) Methyl Acetate	4.31	43	5427	0.830	ug/l #	77
20) Methylene Chloride	4.53	84	11684	2.381	ug/l	96
43) Isopropyl Acetate	8.15	43	123271	10.746	ug/l #	90
80) 1,3,5-Trimethylbenzene	12.73	105	2949	0.209	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101819\
Data File : VN058851.D
Acq On : 18 Oct 2019 12:39
Operator : JC/SP
Sample : PB124110TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB124110TB

Quant Time: Oct 21 07:40:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

