

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101819\
 Data File : VN058852.D
 Acq On : 18 Oct 2019 13:01
 Operator : JC/SP
 Sample : K5439-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 33RD-ST-7TH-AVE-EAST-COMP

Quant Time: Oct 21 07:55:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	394882	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	629572	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	576323	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	200693	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	286930	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
35) Dibromofluoromethane	7.57	113	203892	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
50) Toluene-d8	10.08	98	808985	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
62) 4-Bromofluorobenzene	12.40	95	249731	42.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.64%	
Target Compounds						
16) Acetone	3.80	43	11981	4.186	ug/l	99
20) Methylene Chloride	4.53	84	1217658	249.179	ug/l	100
43) Isopropyl Acetate	8.15	43	91235	8.100	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101819\
Data File : VN058852.D
Acq On : 18 Oct 2019 13:01
Operator : JC/SP
Sample : K5439-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
33RD-ST-7TH-AVE-EAST-COMP

Quant Time: Oct 21 07:55:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

