

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101819\
 Data File : VN058868.D
 Acq On : 18 Oct 2019 19:01
 Operator : JC/SP
 Sample : K5452-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FA19-MMW0423-9001

Quant Time: Oct 21 08:44:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	370271	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	601807	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	536647	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	196204	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	286437	52.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.48%	
35) Dibromofluoromethane	7.57	113	204592	51.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.42%	
50) Toluene-d8	10.08	98	762926	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
62) 4-Bromofluorobenzene	12.40	95	241126	42.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.50%	
Target Compounds						
16) Acetone	3.80	43	9487	3.535	ug/l	95
64) Tetrachloroethene	10.63	164	6453	1.717	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101819\
Data File : VN058868.D
Acq On : 18 Oct 2019 19:01
Operator : JC/SP
Sample : K5452-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FA19-MMW0423-9001

Quant Time: Oct 21 08:44:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

