

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101819\
 Data File : VN058869.D
 Acq On : 18 Oct 2019 19:23
 Operator : JC/SP
 Sample : K5452-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FA19-EX3

Quant Time: Oct 21 08:47:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	375671	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	590765	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	510639	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	175027	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	291853	52.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.94%	
35) Dibromofluoromethane	7.57	113	207575	53.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.88%	
50) Toluene-d8	10.09	98	778245	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
62) 4-Bromofluorobenzene	12.40	95	233239	42.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.24%	
Target Compounds						
3) Chloromethane	2.04	50	1969	0.339	ug/l #	78
4) Vinyl Chloride	2.17	62	5559	0.965	ug/l	93
16) Acetone	3.80	43	17917	6.581	ug/l	96
27) cis-1,2-Dichloroethene	6.81	96	332117	71.085	ug/l	89
44) Trichloroethene	8.82	130	26404	6.103	ug/l	100
64) Tetrachloroethene	10.63	164	47682	13.333	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101819\
Data File : VN058869.D
Acq On : 18 Oct 2019 19:23
Operator : JC/SP
Sample : K5452-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FA19-EX3

Quant Time: Oct 21 08:47:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

