

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN068968.D
 Acq On : 18 Oct 2021 10:31
 Operator : JC/MD
 Sample : VN1018MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1018MBL01

Quant Time: Oct 19 06:32:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	362894	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	634482	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	569019	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	210220	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	240626	50.291	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.580%
35) Dibromofluoromethane	8.024	113	190644	51.644	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.280%
50) Toluene-d8	10.443	98	785952	53.212	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.420%
62) 4-Bromofluorobenzene	12.734	95	264464	47.458	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN068968.D
Acq On : 18 Oct 2021 10:31
Operator : JC/MD
Sample : VN1018MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1018MBL01

Quant Time: Oct 19 06:32:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

